

RESEARCH

Open Access



Loop-mediated isothermal amplification assays for the detection of antimicrobial resistance elements in *Vibrio cholera*

Daniel Antonio Negrón¹, Shipra Trivedi¹, Nicholas Tolli², David Ashford², Gabrielle Melton¹, Stephanie Guertin², Katharine Jennings¹, Bryan D. Necciai³, Shanmuga Sozhamannan^{3,4} and Bradley W. Abramson^{1*}

*Correspondence:
Bradley.Abramson@noblis.org

¹ Noblis, Inc., 2002 Edmund Halley Dr, Reston, VA 20191, USA

² Noblis ESI, 14425 Penrose Pl, Chantilly, VA 20151, USA

³ Joint Program Executive Office for Chemical, Biological, Radiological and Nuclear Defense (JPEO-CBRND), Joint Project Lead for CBRND Enabling Biotechnologies (JPL CBRND EB), Frederick, MD 21702, USA

⁴ Joint Research and Development, Inc., Stafford, VA 22556, USA

Abstract

Background: The bacterium *Vibrio cholerae* causes diarrheal illness and can acquire genetic material leading to multiple drug resistance (MDR). Rapid detection of resistance-conferring mobile genetic elements helps avoid the prescription of ineffective antibiotics for specific strains. Colorimetric loop-mediated isothermal amplification (LAMP) assays provide a rapid and cost-effective means for detection at point-of-care since they do not require specialized equipment, require limited expertise to perform, and can take less than 30 min to perform in resource limited regions. LAMP output is a color change that can be viewed by eye, but it can be difficult to design primer sets, determine target specificity, and interpret subjective color changes.

Methods: We developed an algorithm for the *in silico* design and evaluation of LAMP assays within the open-source PCR Signature Erosion Tool (PSET) and a computer vision application for the quantitative analysis of colorimetric outputs. First, Primer3 calculates LAMP primer sequence candidates with settings based on GC-content optimization. Next, PSET aligns the primer sequences of each assay against large sequence databases to calculate sufficient sequence similarity, coverage, and primer arrangement to the intended taxa, ultimately generating a confusion matrix. Finally, we tested assay candidates in the laboratory against synthetic constructs.

Results: As an example, we generated new LAMP assays targeting drug resistance in *V. cholerae* and evaluated existing ones from the literature based on *in silico* target specificity and in vitro testing. Improvements in the design and testing of LAMP assays, with heightened target specificity and a simple analysis platform, increase utility for in-field applications. Overall, 9 of the 16 tested LAMP assays had positive signal through visual and computer vision-based detection methods developed here. Here we show LAMP assays tested on synthetic AMR gene targets for *aph(6)*, *varG*, *flaR*, *qnrVC5*, and *almG*, which allow for resistance to aminoglycosides, penicillins, carbapenems, phenicols, fluoroquinolones, and polymyxins respectively.

Keywords: LAMP, AMR, Assay, *Vibrio cholera*, Antimicrobial resistance, Image processing, Computer vision, Loop-mediated isothermal amplification, Molecular diagnostic, Colorimetric



Background

Cholera, the diarrheal illness caused by ingestion of *Vibrio cholerae*, affects 1.3–4 million people worldwide, where 1 in 10 develop severe symptoms caused by dehydration according to the US Center for Disease Control and Prevention (CDC) [1]. In 2022, 80 countries reported data on cholera to the World Health Organization (WHO). Of these, 44 countries reported 472,697 cases and 2,349 deaths, a case-fatality rate (CFR) of 0.5% in both outbreaks and as imported cases [2]. Prevention and treatment options are available, such as the oral cholera vaccines (OCVs) and intravenous rehydration therapy, which, together with good hygiene practices, are optimal for reducing spread, morbidity, and mortality [3, 4]. However, due to vaccine shortages, antibiotics may still be recommended to help reduce the severity and duration of symptoms, especially in severe or special cases, such as pregnant individuals or those with comorbidities. Unfortunately, widespread antibiotic use has led to increasing reports of antimicrobial resistance (AMR) and multidrug resistance (MDR) in *V. cholerae* [5–7]. To slow the acquisition of further resistance among cholera strains, prophylactic antibiotic treatment is not advised. Therefore, targeting the prescribed antibiotic as precisely as possible is ideal.

Frontline medical providers can optimize medical treatments based on point-of-care diagnostics, which can reduce lengthy analysis times and help improve patient outcomes. For example, molecular diagnostics such as polymerase chain reaction (PCR) and quantitative PCR (qPCR) assays quickly and reliably amplify target genetic material to detect the presence of targeted pathogenic organisms [8]. However, field usage is limited since they require bulky or expensive thermocycler instruments. Recently, loop-mediated isothermal amplification (LAMP) assays have become popular for bacterial and viral sample analyses; they are quicker than PCR, do not require expensive equipment, and provide a visual colorimetric readout of positive detection [9–13].

However, LAMP assays present challenges that slow adoption and limit their usefulness. LAMP assays require up to six primers in complex orientations to generate the required loop structure and can lead to inconsistencies in the design of primer sequences and spacing but still produce robust results compared to traditional methods [14]. However, with a greater number of primers compared to traditional PCR and qPCR, there is a greater chance for primer dimers, mispriming, and false positive. Primer panels for LAMP assays are more difficult to design than PCR and qPCR assays, since the primer set solution space is constrained by additional thermodynamic, length, and distance requirements. Several software applications, like the NEB® LAMP Primer Design Tool or Primer Explorer, can help design primers based on user-input target sequences, but these applications do not provide specificity metrics for the resulting assay with respect to taxonomy [15–17]. Second, once a successful assay panel has been designed, the resulting colorimetric change must be assessed visually for the presence or absence of the pathogen. Subjective human interpretation of color can be unreliable, time-consuming, and limited in sensitivity [9, 18]. To address this need, recently, visible colorimetric measurements using smartphone or tablet software and cameras (i.e., smartphones) have been developed [19–21].

The PCR Signature Erosion Tool (PSET) provides an *in silico* method for determining primer binding sites relative to taxonomically related sequences and detects mutations in primers used for PCR or qPCR assays [22, 23]. For this study, we created LAMP assays

for *V. cholerae* detection, beginning with improvements to PSET, and culminating with a robust, quantitative determination of the assay outcome using computer vision. The new LAMP workflow module for PSET can generate assays or predict performance using large reference sequence databases, such as the NCBI nucleotide (nt) database. Running the tool periodically can identify potential assay failures as newly evolved sequences are deposited into databases such as GenBank or GISAID [22]. Additionally, we developed a LAMP assay design tool based on the open source Primer3 program as an alternative to web-based programs, leveraging multicore architecture. We then performed end-to-end design of novel LAMP assays together with *in silico* evaluation of existing ones. Finally, we validated the LAMP assay designs *in vitro* and developed a method to computationally determine the colorimetric changes of the LAMP assay using a smartphone camera. The code is publicly available at <https://github.com/biolaboro/PSET>.

Methods

Design *in silico*

A LAMP assay design script was written in Python (v3.11.4) with bindings to the Primer3 PCR design tool (v2.6.1) via the primer3-py (v2.6.0) API [24–27] and a schematic representation provided in Figure S2. A Snakemake workflow coordinates the execution of the script, which first runs Primer3 to identify primer pairs and optional loop candidates [28]. It uses the Primer3 configuration file system to guide the nested search for inner primers and loops with additional parameters for inter-primer distances, strandedness, and thermodynamic constraints. LAMP assays were designed using global parameters to space primers (DIST_RANGE) and calculate thermodynamic properties as: F1cB1c_DIST_RANGE=1–2000, F2F1c_DIST_RANGE=20–80, F3F2_DIST_RANGE=1–50, dv=0.5 (divalent cation concentration in mM), mv=200 (monovalent cation concentration in mM), with additional primer-pair specific parameters shown in Table S8. A procedure then evaluates every primer pair and loop primer combination. Accordingly, there is an initial search for the outer F3/B3 primers and inner F2/B2 primers. The next four searches calculate the F1c, B1c, LF, and LB candidates separately. Nonoverlapping combinations of F2-LF-F1c and B1c-LB-B2 are then evaluated based on F1c-B1c distance constraints and for 5'/3' stability in terms of ΔG . The final LAMP 5'/3' assay definition indicates primers and loops with square brackets and parentheses on the amplicon. Additionally, each primer, loop, FIP, and BIP component is individually defined in a resulting JSON object with an associated assay penalty score equal to the sum of Primer3 penalties.

Evaluation *in silico*

An alignment-based workflow implements the LAMP assay evaluation within the PSET framework. The initial phase queries the amplicon against a BLAST+ database [29], collecting a list of accessions with sequence identity and coverage above a parameterized threshold. The next phase realigns each component primer and loop separately, using the glsearch36 tool of the FASTA suite [30] for a global–local alignment to guarantee full query coverage. A filtering step removes results with alignment below a parameterized identity threshold. The last phase checks for nested primer and loop order, arrangement, and inter-primer/loop distances. A confusion matrix is output with respect to each

subject accession, where calls are based on subject taxonomy and whether primer and loop sequences are arranged correctly while exceeding identity threshold. The taxonomy check determines whether the NCBI Taxonomy identifier of the subject is equal to or a descendant of any of the identifiers in the set of assay targets. Accordingly, the taxonomy and arrangement-alignment evaluation yields the true/false and positive/negative component of the confusion matrix call.

In vitro LAMP assay testing

All LAMP primers (Table S1) were ordered from IDT and reconstituted at 100 μ M in sterile water. FIP and BIP primers were diluted to 16 μ M, F3 and B3 primers were diluted to 2 μ M, and LF and LB primers were diluted to 4 μ M to make a 10X primer master mix. Positive control gene Block (gBlock) fragments were ordered from IDT and reconstituted at 10ng/ μ L. Serial (1:10) dilutions were performed in molecular grade water to achieve the concentrations of 10pg/ μ L through 1fg/ μ L that were tested. LAMP reagents were ordered from NEB (WarmStart Colorimetric LAMP 2X Master Mix cat# M1800S). The LAMP reaction mix was prepared using 12.5 μ L of WarmStart Colorimetric LAMP 2X Master Mix, 2.5 μ L of the reaction specific LAMP 10X primer mix, and 9 μ L of molecular grade water per well of a 96-well reaction plate (Applied Biosystems cat# N8010560). The final primer concentrations were 1.6 μ M for FIP and BIP, 0.2 μ M for F3 and B3, and 0.4 μ M for LF and LB. 1 μ L of the desired dilution of each gene fragment was then added to the corresponding well and 1 μ L of molecular grade water was added to no template control wells. The plate was incubated at 65°C for 30 min in a thermocycler (Applied Biosystems TFS-2720) following LAMP 2X Master Mix manufacture guidelines. After the incubation period, the plate was allowed to cool to room temperature. Final LAMP assay images were captured with a smartphone (Samsung Galaxy S10e) with default camera settings.

LAMPvision image analysis

The LAMPvision prototype python script uses the NetworkX (3.0), NumPy (v1.26.4), OpenCV (v4.9.0), Pandas (2.2.0), Scikit-Learn (1.5.0) packages to analyze the LAMP assay images [31–35]. First, the algorithm filters the image to restrict colors within the HSV (hue, saturation, value) color range between (0, 85, 100) and (179, 255, 255), which correspond to pinkish reds and yellows of the LAMP assay. Next, the Hough circle transform procedure detects circles within the grayscale representation of the image. This step returns the location and size of each circle detected and is meant to locate plate wells. Overlapping circles are combined into a bigger one based on radius and the new location is set to the centroid. Next, the circles are binned into a grid index based on horizontal and vertical overlap. Accordingly, a 1-D clustering step uses the mean-shift algorithm to calculate the minimum cluster value based on the pairwise distances of the circles. This requires that the well plate have at least two non-empty adjacent wells. The resulting value reflects the average distance between adjacent wells. This is done so that the output corresponds to each well plate index, even if a row or column is empty. Values for color and circle detection were determined through trial and error and suitability on these images and values are hardcoded into the LAMPvision script in the GitHub repository. Finally, the mean RGB color of each circle is output with the corresponding well

index. An R script (v4.4.0) calculated the color difference between each reaction and the corresponding no-template negative control and rendered plots with the ggplot2 library (3.5.1) and related extensions [36–42]. These color differences across replicates (n=4 for each assay sample) were then plotted with the relative copies per well. Uniquely, the graph plots each point as the extracted well pixels from the original image. Raw images, metadata, and scripts are available in the analysis.zip archive file.

Results

LAMP assay primer design and PSET analysis of *V. cholerae* AMR genes

A functional LAMP assay requires careful primer design, with considerations such as amplicon length, primer spacing, primer orientation, and primer annealing temperatures, while also requiring specificity to the target sequence [11]. Here we designed and incorporated a LAMP primer design tool into PSET for rapid and automated determination of primer erosion that occurs when mutations in the template strand mismatch with the primer, causing failure of a previously working assay (Fig. 1). Briefly, the LAMP assay design algorithm uses Primer3 to determine the F3/B3 primer pair, followed by the internal primer pairs F2/F1c and B1c/B2 which are concatenated into FIP and BIP,




Fig. 1 Conceptual design of LAMP primers. **a** The user loads assays as LAMP primers or in an assay definition format for **b** blasting to find target sequences and validate primer spacing. Appropriate primer alignments, primer arrangements, and taxonomy targets are calculated. **c** A confusion matrix is generated for each assay to describe the target specificity

respectively. Within the latter pairs, additional optional primers are chosen to become Loop-F (LF) and Loop-B (LB) primers, which can help speed up nucleic acid amplification during the assay. The final LAMP assay is designed based on the appropriate spacing of these primer sets as well as the annealing temperature differences to create a constraint-based LAMP assay for the target sequence (see Methods for detailed parameters).

PSET evaluates LAMP assays by analyzing primer alignment and arrangement against BLAST database taxa to determine *in silico* target specificity. The previous version of PSET determined PCR or qPCR assays which have 2 or 3 primer sets respectively. Therefore, an algorithm was required for LAMP assays containing 6–8 total primers which are not in a typical PCR primer orientation (Fig. 1). Since the FIP/BIP primers create a loop structure through an inverted hairpin, their sequences are a combination of disjointed forward and reverse primers F2/F1c or B2/B1c, respectively. An algorithm was developed to reduce the FIP and BIP primers to their counterparts. Briefly, a BLAST searches queries for alignment of both F3 and B3 primers to retrieve a template sequence from the database and extracts a complete sequence template for determining the location of the other primer sequences. Then FIP/BIP are aligned to the template with BLAST. Each FIP and BIP primer should produce two alignments where one alignment is forward, and one alignment is reverse. The hits are then evaluated where the FIP is expected to have at least one alignment in the forward direction and one alignment in the reverse direction where the sum of the two alignment lengths should equal the length of the original FIP primer. The location of the F3/B3, F2/F1c, and B2/B1c relative to the template are used for evaluation of primer spacing. If F2/F1c, and B2/B1c are not found in the target sequence then an additional query is used as a template.

For integration into the PSET evaluation routine, LAMP assays are treated as three pseudo-PCR/qPCR assays representing amplicons from the pairs of primer for F3/B3, F2/F1c with LF denoted as a probe, and B2/B1c with LB denoted as a probe. Each one is evaluated for correct orientation and spacing. The final output is a cumulative “call” of the LAMP assays that reflect primer alignment, arrangement, and subject sequence taxonomy. True Positive (TP) LAMP assay targets must have all pseudo-PCR/qPCR assays return individual TP calls to the same subject. If any pseudo-PCR/qPCR assays return a False Negative (FN) then the LAMP assay is considered an FN, even if primers are oriented correctly, which demonstrates that the overall LAMP assay would not effectively target the subject sequence of interest.

We demonstrated PSET’s taxonomy detection feature using previously published PCR, qPCR, or LAMP primer sets targeting *V. cholerae* AMR genes, while additionally designing novel LAMP primer sets (Table S1). These primer sets were compared to NCBI’s nt database targeting NCBI Taxonomy identifier 666 (*V. cholerae*) using PSET (Table S2). In general, previously published assays largely targeted *V. cholerae*, however the nt database does not provide a means of distinguishing submissions with AMR. Therefore, some of the AMR targets like *gyrA*, *tcp*, and *ompU* genes result in high FP hits, suggesting that greater than 80% of the submitted nucleotide sequences do not contain these specific AMR genes (Table S2). Therefore, we created a custom taxonomy database using *V. cholerae* sequences from Microbial Browser for Identification of Genetic and Genomic Elements (MicroBIGG-E) that are associated with AMR classes and genes. Given that PSET

operates using NCBI's taxonomy database, we created a system to import new taxonomies based on AMR classes from MicroBIGG-E (Table S3) and tested the LAMP assays designed here (Table S4). 34 of 43 assays showed an AMR class specificity of greater than 97%. However, the assays with low specificity targeted tetracycline and quinolone where these AMR classes have multiple genes for conferring resistance. This is likely due to gene-specific targeting since several are capable of conferencing resistance to the same class.

In vitro testing with independent quantitative analysis methods

Primers were designed using the design principles described above and the target gene fragments were synthesized based on deposited sequences from MicroBIGG-E [43]. The samples were serially diluted, producing a tenfold dilution curve for each assay alongside a sample that did not contain any template as a negative control (Fig. 2). Four independent replicates were incubated for 30min at 60 °C. These LAMP colorimetric assays produce a pH dependent color shift in response to nucleic acid amplification where negative samples are red and positive samples are yellow. A quantitative visual interpretation of the color change was determined by two individuals by eye (Table S5) as well as an automated interpretation based on image capture and hexadecimal-valued color distance




Fig. 2 In vitro LAMP assay analysis. **a** Plate layout scheme for the replicates (n=4) performed. Each assay consisted of 5 samples with template and a no template control. A total of 16 unique assays were run on one plate and performed four separate days. **b** Image of the completed LAMP assay with detected wells within the green circles. **c** Extracted hexadecimal color values. **d** Each sample's color difference to the no template control was calculated. **e** One assay displayed zoomed in describing dilutions and copies of the template used. **f** The color difference was plotted relative to the copies of the template for each reaction for the four replicates. Each subplot includes a linear model with the formula and associated statistics computed with the stat_poly_eq function of the ggpmisc R package, including the R^2 , F value, and model fit p-value. A table of these and additional statistics are provided in Table S9

(Table S6 and S7). Nine of the 16 assays produced positive signals in the highest concentration template reaction while the no-template control was still negative, and four assays remained negative at all DNA template concentrations suggesting no amplification occurred. Three of the assays had amplification in all reactions, including the negative control, and further testing is required to confirm if this possibly from primer dimer amplification. Based on the biological replicates the authors do not hypothesize this to be contamination.

Visual analysis of LAMP assays showed detection of five AMR gene targets for *aph(6)*, *varG*, *floR*, *qnrVC5*, and *almG*, which allow for resistance to aminoglycoside antibiotic, penicillins and carbapenems, phenicol antibiotics, fluoroquinolone, and polymyxin respectively. Analysis by visual interpretation is subjective, time consuming, and difficult to interpret individual wells on a plate [44]. Therefore, automated image analysis was employed. Using a cellphone to capture images of the plates, we determined color differences of the sample to their cognizant no-template control (Fig. 2d). These color differences across replicates ($n=4$ for each reaction) are then plotted with the relative copies per well (Fig. 2f and S1). Each point on the graph consists of the extracted well pixels from the original image.

The slope was calculated from the LAMPvision color distance plots to determine if assays have a positive signal trajectory with increasing template concentration and it was found that a slope threshold of 4.0 may be used to determine if positive signal is produced, but further testing is required (Fig. 3a). Next, we determined the correlation of the methods for the LAMP assay visual analysis and the LAMPvision tool (Fig. 3). Eight assays showed a high degree of correlation (>0.8 Pearson correlation) between LAMP vision and the visual quantitative analysis (Fig. 3a). Correlations were not calculated for four assays because the visual analysis was not able to determine any positive signal and all values were zero by visual analysis (Table S5). Three assays did not show a color change but had a strong correlation (0.5–0.8 Pearson correlation) and one assay had a negative correlation. These results suggest there is a strong agreement in the visual analysis and LAMPvision analysis when there is positive signal of the assay.

Discussion


Here we show the in vitro application of *in silico* LAMP assay design incorporated into the Primer Signature Erosion Tool (PSET) and a computer vision algorithm to interpret complex plate-based colorimetric LAMP assays. By combining all these aspects, we can rapidly design, test, and validate novel LAMP assays targeting AMR genes while also determining the suitability of previously published LAMP assays for target specificity. The rapid design and validation of these LAMP assays allow individuals to develop vast arrays of assays targeting numerous genes, signature sequences of species, or variants of concern. Here we validated the code to produce novel LAMP assays and detect AMR synthetic gene sequences. Although this work primarily focused on in silico aspects, limited in vitro data show great promise for on-demand assay design and testing in far-forward scenarios, applications, and use-cases.

Colorimetric LAMP assays as a diagnostic have been approved by the FDA for some target strains, such as *Salmonella* [45] or SARS-CoV-2 [46], via Emergency Use Authorization (EUA) which demonstrates their utility. The reliability of a nucleic acid detection


a. Correlation of human visual vs quantitative color determination

Assay Number	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
Slope	7.85	8.79	11.24	5.59	4.34	3.08	0.90	1.20	0.97	10.16	8.43	8.51	3.14	0.91	4.10	8.63
LAMPVision vs Eye (Pearson)	0.93	0.95	0.96	0.65	NC	NC	-0.32	NC	NC	0.99	0.98	0.97	0.67	0.55	0.87	0.98
Eye1 vs Eye2 (Pearson)	1.00	0.99	1.00	1.00	NC	NC	0.88	NC	NC	1.00	1.00	1.00	0.92	0.67	1.00	0.99

b. Human vs human



d. Example color detection by assay & method



c. Humans vs computer




Fig. 3 Comparison of visual interpretation to quantitative computer LAMPvision methods. **a** The Pearson correlation was calculated for the human visual determination of color shift on a scale of 0–5 and compared to the LAMPvision methods described here. NC indicates non-computable due to standard deviation of 0. **b** A comparison describing the subjective nature of two human individuals’ interpretation of the color change. **c** Another comparison of color scores by visual analysis (Eye1, top; Eye2, bottom) versus computer vision analysis. **d** A plot showing the difference in color change detected visually and by computer vision methods. Computer vision methods enhance the limit of detection. A table of additional statistics are provided in Table S9 and Table S10 for subfigures a, b, and c

method that is easy to use and interpret will help track novel and recurring pathogens, aiding in the surveillance of non-infectious environmental DNA (eDNA), responding to epidemic scenarios, and allowing for diagnostic use in low-resource environments. Ussing the software designed here, new LAMP assays can be designed, analyzed, and output into an ordering form for rapid prototyping as demonstrated here (Fig. 2). These methods produce standard LAMP primer sets that should feed into existing regulatory workflows; however, proper clinical sample testing needs to be performed.

In silico target specificity was calculated for 66 previously published LAMP, PCR, and qPCR assays from 27 publications targeting AMR genes in *V. cholerae* (Table S1). Additionally, we designed 43 new LAMP assays with the described LAMP primer design and PSET integration tools. These assays were tested via PSET for their *in silico* target specificity against NCBI’s nt database (Table S2) and the *V. cholerae* subset of the Micro-BIGG-E database (Table S3).

LAMP colorimetric assays are used as rapid, point-of-care diagnostics. However, due to the non-specific pH-based color changes, they cannot be multiplexed in a single well. An array of individual colorimetric LAMP assays can be used to detect multiple targets, but then the interpretation of numerous assays becomes problematic. Here we designed a useful computer vision tool, LAMPvision, to detect wells and quantify results reliably across several replicates ($n=4$). Other studies have used wavelength shifts in absorption spectra or smartphone applications, but those studies required expensive machines or complicated setups to correct for lighting [45, 47]. LAMPvision uses an internal negative control placed next to the samples to serve as a color normalization factor. However, there were observable differences in saturation across the images that may introduce signal noise that were not corrected for in this study. These may be mitigated by performing additional background correction processes, like nonuniform nonlinear brightness corrections which account for changes in brightness on the edges of the image relative to the focal point. Although lighting standards could be employed, the aim of these computer vision methods are to reduce cognitive load and use automated image correction to make assays easier to interpret. All these methods require additional testing with different lighting color conditions and different smart phone camera types.

Several of the LAMP assays were not functional and required revisions to the ordering form of the code, initially. For instance, the B3 primer was ordered in the wrong orientation relative to the assay due to a bug in the output script. However, the assays were still functional and produced positive signals. Interestingly, the utility of the B3 primer is to facilitate the removal of the FIP/BIP primers during amplification. This could be one reason these assays have high detection limits regarding copy number. Further refinement of these metrics will presumably result in a higher rate of LAMP assay success. Similarly, since only synthetic DNA controls were used in these experiments, further characterization on live samples is required. Furthermore, much more stringent controls and validation steps are required for clinical usage.

Color image analysis is difficult to compare to human eye perception and additional testing may be required to validate the color distance calculations [44]. In this method, we used red, green, and blue color values which do not account for color intensity and perhaps a color difference calculation using hue, saturation, lightness (value; HSL, HSV) is better. Since RGB hexadecimal values can fluctuate with brightness, intensity correction methods may be required for validation and could be implemented using negative control wells strategically interspersed as a color correction factor. Here, we did not determine a specific negative control color value for baseline and instead used the associated assays' no-template control as a baseline value for a color change distance measure. Accordingly, having one negative control does not provide a means to calculate negative standard deviations for thresholding positive signals. In future applications, all sample negative controls may be utilized but would require brightness and color correction across the plate and across separate images.

Future improvements to these methods would result in a greater percentage of functional in vitro assay successes utilizing clinical samples. A major limitation to this study is the use of synthetic controls, but clinical samples will be utilized in future experiments. Likewise, careful examination of the very strict thermodynamic properties of each primer from published LAMP primer sets will help refine and constrain primer

attributes leading to greater in vivo success, thereby limiting potential primer dimer based false positives. Finally, sequencing the LAMPlicon product may be beneficial in determining if the false positive events are the result of DNA amplification or some other pH change causing the color shift.

Conclusion

Overall, we present a means of rapidly prototyping new LAMP assays with target specificity including in vivo tests with synthetic DNA sequences with computer vision analysis. LAMP's point-of-care usage can help combat increasing antimicrobial resistance as LAMP assays become easier to access and interpret for individuals and clinicians. As these tests become robust and well validated, they will have greater utility in field testing to control and mitigate cholera outbreaks. As outlined in the Global Roadmap 2030.2, intensified strategic and systematic use of rapid diagnostic tests followed by laboratory confirmation will contribute to reaching the targets of the “Ending Cholera: Ending Cholera: a Global Roadmap to 2030” initiative [2].

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s12859-024-06001-3>.

Additional file 1.

Additional file 2.

Additional file 3.

Acknowledgements

The views expressed in this manuscript are those of the authors and do not necessarily reflect the official policy or position of the JPEO-CBRND, the Department of Defense, or the U.S. Government. This work was prepared as part of the author(s) official duties. Title 17 U.S.C. § 105 provides that ‘Copyright protection under this title is not available for any work of the United States Government.’ Title 17 U.S.C. §101 defines U.S. Government work as work prepared by a military service member or employee of the U.S. Government as part of that person's official duties. References to non-federal entities or their products do not constitute or imply Department of Defense or Army endorsement of any company, product or organization. Funding for this study was provided and executed by the Joint Program Executive Office for Chemical, Biological, Radiological and Nuclear Defense's (JPEO-CBRND) Joint Project Lead for CBRND Enabling Biotechnologies (JPL CBRND EB) on behalf of the Department of Defense's Chemical and Biological Defense Program.

Author contributions

Conceptualization: DN, DA, SS, and BA. Methodology: DN, ST, SS, and BA. Software: DN, ST, and BA. Validation: DN, NT, GM, and BA. Formal analysis: DN, and BA. Investigation: DN, NT, GM, and BA. Resources: KJ, SS, and BA. Data Curation: DN, and BA. Writing—Original Draft: DN, NT, and BA. Writing—Review & Editing: DN, ST, NT, DA, GM, SG, KJ, BN, SS, and BA. Visualization: DN, and BA. Supervision: KJ, SS, and BA. Project administration: DN, DA, KJ, BN, SS, and BA. Funding acquisition: KJ, BN, SS, and BA.

Funding

Funding for this work was provided by JPEO (CBRND), JPL-EB, DBPAO (contract number W911SR-22-C-0049) and laboratory work was funded by the Noblis Sponsored Research program for internal R&D.

Availability of data and materials

The datasets analyzed during the current study are available in the NCBI BLAST (<https://ftp.ncbi.nlm.nih.gov/blast/db/v5/>), NCBI Taxonomy (<https://ftp.ncbi.nlm.nih.gov/pub/taxonomy/>), and MicroBIGG-E (<https://www.ncbi.nlm.nih.gov/pathogens/microbigge>) repositories. Data generated during this study are included in this published article and its supplementary information files. Code is publicly available at <https://github.com/biolaboro/PSET>.

Declarations

Ethics approval and consent to participate

Not applicable.

Consent for publication

Not applicable.

Competing of interests

Not applicable.

Received: 16 September 2024 Accepted: 26 November 2024

Published online: 18 December 2024

References

- Ali M, Nelson AR, Lopez AL, Sack DA. Updated global burden of cholera in endemic countries. *PLoS Negl Trop Dis*. 2015;9(6):e0003832.
- WHO. Cholera, 2022. WER. 2023; 98(38): 431–43.
- WHO. Cholera vaccines: WHO position paper–August 2017. WER. 2017;92(34): 477–500
- CDC. Cholera. [cited 2024 May 21]. Treating Cholera. 2024 Available from: <https://www.cdc.gov/cholera/treatment/index.html>
- Das B, Verma J, Kumar P, Ghosh A, Ramamurthy T. Antibiotic resistance in *Vibrio cholerae*: understanding the ecology of resistance genes and mechanisms. *Vaccine*. 2020;29(38):A83–92.
- Yuan X-H, Li Y-M, Vaziri AZ, Kaviar VH, Yang Jin Yu, Jin AM, et al. Global status of antimicrobial resistance among environmental isolates of *Vibrio cholerae* O1/O139: a systematic review and meta-analysis. *Antimicrob Resist Infect Control*. 2022;11(1):62.
- Verma J, Bag S, Saha B, Kumar P, Ghosh TS, Dayal M, et al. Genomic plasticity associated with antimicrobial resistance in *Vibrio cholerae*. *Proc Nat Academy Sci*. 2019;116(13):6226–31.
- Dutta D, Naiyer S, Mansuri S, Soni N, Singh V, Bhat KH, et al. COVID-19 diagnosis: a comprehensive review of the RT-qPCR method for detection of SARS-CoV-2. *Diagnostics*. 2022;12(6):1503.
- Kashir J, Yaqinuddin A. Loop mediated isothermal amplification (LAMP) assays as a rapid diagnostic for COVID-19. *Med Hypotheses*. 2020;1(141): 109786.
- Zhao X, Zeng Y, Yan B, Liu Y, Qian Y, Zhu A, et al. A novel extraction-free dual HiFi-LAMP assay for detection of methicillin-sensitive and methicillin-resistant *Staphylococcus aureus*. *Microbiol Spectrum*. 2024;12(4):e04133–e4223.
- Notomi T, Mori Y, Tomita N, Kanda H. Loop-mediated isothermal amplification (LAMP): principle, features, and future prospects. *J Microbiol*. 2015;53(1):1–5.
- Bao H, Zhao Y, Wang Y, Xu X, Shi J, Zeng X, et al. Development of a reverse transcription loop-mediated isothermal amplification method for the rapid detection of subtype H7N9 avian influenza virus. *Biomed Res Int*. 2014;2014(6): e525064.
- Leroy AG, Persyn E, Gibaud SA, Crémet L, Le Turnier P, Benhamida M, et al. Assessment of a multiplex LAMP assay (Eazyplex® CSF Direct M) for rapid molecular diagnosis of bacterial meningitis: accuracy and pitfalls. *Microorganisms*. 2021;9(9):1859.
- Francois P, Tangomo M, Hibbs J, Bonetti EJ, Boehme CC, Notomi T, et al. Robustness of a loop-mediated isothermal amplification reaction for diagnostic applications. *FEMS Immunol Med Microbiol*. 2011;62(1):41–8.
- New England Biolabs. NEB® LAMP Primer Design Tool [Internet]. [cited 2024 May 24]. Available from: <https://lamp.neb.com/#/>
- Eiken Chemical Co., Ltd. LAMP primer designing software PrimerExplorer [Internet]. [cited 2024 May 24]. Available from: <https://primerexplorer.jp/e/index.html>
- Akhmetzianova LU, Davletkulov TM, Sakhabutdinova AR, Chemeris AV, Gubaydullin IM, Garafutdinov RR. LAMPprimers iQ: new primer design software for loop-mediated isothermal amplification (LAMP). *Anal Biochem*. 2024;1(684): 115376.
- Ahn SJ, Baek YH, Lloren KKS, Choi WS, Jeong JH, Antigua KJC, et al. Rapid and simple colorimetric detection of multiple influenza viruses infecting humans using a reverse transcriptional loop-mediated isothermal amplification (RT-LAMP) diagnostic platform. *BMC Infect Dis*. 2019;19(1):676.
- Colbert AJ, Lee DH, Clayton KN, Wereley ST, Linnes JC, Kinzer-Ursem TL. PD-LAMP smartphone detection of SARS-CoV-2 on chip. *Anal Chim Acta*. 2022;22(1203): 339702.
- García-Bernalt Diego J, Fernández-Soto P, Márquez-Sánchez S, Santos Santos D, Febrer-Sendra B, Crego-Vicente B, et al. SMART-LAMP: a smartphone-operated handheld device for real-time colorimetric point-of-care diagnosis of infectious diseases via loop-mediated isothermal amplification. *Biosensors*. 2022;12(6):424.
- Heithoff DM, Barnes LV, Mahan SP, Fox GN, Arn KE, Ettinger SJ, et al. Assessment of a smartphone-based loop-mediated isothermal amplification assay for detection of SARS-CoV-2 and influenza viruses. *JAMA Netw Open*. 2022;5(1):e2145669.
- Negrón DA, Kang J, Mitchell S, Holland MY, Wist S, Voss J, et al. Impact of SARS-CoV-2 Mutations on PCR Assay Sequence Alignment. *Frontiers in Public Health* [Internet]. 2022 [cited 2022 Apr 28];10. Available from: <https://www.frontiersin.org/article/10.3389/fpubh.2022.889973>
- Sozhamannan S, Holland MY, Hall AT, Negrón DA, Ivancich M, Koehler JW, et al. Evaluation of signature erosion in Ebola virus due to genomic drift and its impact on the performance of diagnostic assays. *Viruses*. 2015;7(6):3130–54.
- Untergasser A, Cutcutache I, Koressaar T, Ye J, Faircloth BC, Remm M, et al. Primer3: new capabilities and interfaces. *Nucleic Acids Res*. 2012;40(15):e115.
- Koressaar T, Remm M. Enhancements and modifications of primer design program Primer3. *Bioinformatics*. 2007;23(10):1289–91.
- Koressaar T, Lepamets M, Kaplinski L, Raime K, Andreson R, Remm M. Primer3_masker: integrating masking of template sequence with primer design software. *Bioinformatics*. 2018;34(11):1937–8.
- Van Rossum G, Drake FL. Python 3 reference manual. Scotts Valley, CA: CreateSpace. 2009.
- Mölder F, Jablonski KP, Letcher B, Hall MB, Tomkins-Tinch CH, Sochat V, et al. Sustainable data analysis with Snake-make [Internet]. F1000Research; 2021 [cited 2023 May 17]. Available from: <https://f1000research.com/articles/10-33>

29. Camacho C, Coulouris G, Avagyan V, Ma N, Papadopoulos J, Bealer K, et al. BLAST+: architecture and applications. *BMC Bioinform.* 2009;10(1):421.
30. Pearson WR. FASTA Search Programs. In: John Wiley & Sons Ltd, editor. *eLS* [Internet]. Chichester, UK: John Wiley & Sons, Ltd; 2014 [cited 2020 May 11]. p. a0005255.pub2. Available from: <http://doi.wiley.com/10.1002/9780470015902.a0005255.pub2>
31. Hagberg A, Swart PJ, Schult DA. Exploring network structure, dynamics, and function using NetworkX [Internet]. Los Alamos National Laboratory (LANL), Los Alamos, NM (United States); 2008 [cited 2024 May 22]. Report No.: LA-UR-08-05495; LA-UR-08-5495. Available from: <https://www.osti.gov/biblio/960616>
32. Harris CR, Millman KJ, Van Der Walt SJ, Gommers R, Virtanen P, Cournapeau D, et al. Array programming with NumPy. *Nature.* 2020;585(7825):357–62.
33. The pandas development team. *pandas-dev/pandas: Pandas* [Internet]. Zenodo; 2020. Available from: <https://doi.org/10.5281/zenodo.3509134>
34. McKinney W. Data Structures for Statistical Computing in Python. In: Walt S van der, Millman J, editors. *Proceedings of the 9th Python in Science Conference.* 2010. 56–61.
35. Pedregosa F, Varoquaux G, Gramfort A, Michel V, Thirion B, Grisel O, et al. Scikit-learn: machine learning in python. *J Mach Learn Res.* 2011;12(85):2825–30.
36. R Core Team. *R: A language and environment for statistical computing* [Internet]. Vienna, Austria. 2020. Available from: <https://www.R-project.org/>
37. Wickham H. *ggplot2: Elegant graphics for data analysis* [Internet]. Springer-Verlag New York. 2016. Available from: <https://ggplot2.tidyverse.org>
38. Yu G. *ggimage: Use Image in "ggplot2"* [Internet]. 2023. Available from: <https://CRAN.R-project.org/package=ggimage>
39. Aphalo PJ. *ggpmisc: Miscellaneous Extensions to "ggplot2"* [Internet]. 2024. Available from: <https://CRAN.R-project.org/package=ggpmisc>
40. Campitelli E. *ggnewscale: Multiple Fill and Colour Scales in "ggplot2"* [Internet]. 2024. Available from: <https://CRAN.R-project.org/package=ggnewscale>
41. Robinson D, Hayes A, Couch S. *broom: Convert Statistical Objects into Tidy Tibbles* [Internet]. 2024. Available from: <https://CRAN.R-project.org/package=broom>
42. Aphalo PJ. *ggpp: Grammar Extensions to "ggplot2"* [Internet]. 2024. Available from: <https://CRAN.R-project.org/package=ggpp>
43. Feldgarden M, Brover V, Fedorov B, Haft DH, Prasad AB, Klimke W. Curation of the AMRFinderPlus databases: applications, functionality and impact. *Microbial Genomics* [Internet]. 2022 [cited 2024 Sep 16]: Available from: <https://www.microbiologyresearch.org/content/journal/mgen/10.1099/mgen.0.000832>
44. de Oliveira CB, Sanchuki HBS, Zanette DL, Nardin JM, Morales HMP, Fornazari B, et al. Essential properties and pitfalls of colorimetric Reverse transcription loop-mediated isothermal amplification as a point-of-care test for SARS-CoV-2 diagnosis. *Molecular Med.* 2021;27(1):30.
45. FDA. Confirmation of Salmonella Isolates by Loop-Mediated Isothermal Amplification (LAMP).
46. Food, Administration D, others. Color SARS-CoV-2 RT-LAMP Diagnostic Assay-EUA Summary. Vol. 252021 [Internet]. Available from: <https://www.fda.gov/media/138249/download>
47. Papadakis G, Pantazis AK, Fikas N, Chatziioannidou S, Tsiakalou V, Michaelidou K, et al. Portable real-time colorimetric LAMP-device for rapid quantitative detection of nucleic acids in crude samples. *Sci Rep.* 2022;12(1):3775.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.